



# Clavicle Plate


Clavicle Fixation System






## Exposure and Reduction

- Center the incision over the fracture site.
- Create and mobilize full thickness flaps taking care to avoid and protect the supraclavicular nerves.
- Expose the clavicle between the deltoid-trapezius interval.
- Reduce the fracture and provisionally hold the fragments.




## Position Plate

- Select a superior or anterior plate that accommodates the fracture and fits the bone.
- Clavicle Plates are equipped with an assortment of screw holes (2.7mm and 3.2mm): Locking, Slotted Glide, and Slotted Compression.
- Consider the hole type and position in relation to the fracture when selecting and placing the plate.
- **Non-locking** screws can be used in any hole.




## Secure Plate

- Use 2.3mm (red) drill for 3.2 screws.
  - Use GUIDELCBS-2.3 for locking screws.
  - Use GUIDE-2.3/3.2 for non-locking screws.
- Use 2.0mm (yellow) drill for 2.7 screws.
  - Use GUIDELCBS-2.0 for locking screws.
  - Use GUIDE-2.0/2.7 for non-locking screws.




## Compress Fracture

- Insert a 3.2 cortical screw into the end of the compression hole away from the fracture and tighten.



## Capture/Reduce Fragments

- Insert a 3.2 cortical screw into the center of the compression hole at an angle that best captures fragments.



## Final Fixation

- Complete fixation with additional screws.

All implants made from surgical grade stainless steel

### Superior Midshaft

CPSMIDx-7  
CPSMIDx-8  
CPSMLTx-8  
CPSMIDx-9  
CPSMIDx-10

x=Left or Right



### Superior Lateral

CPSLATx-11  
CPSLATx-12  
CPSLATx-15\*

x=Left or Right

\*Special Order



### Anterior Midshaft

CPAMID-8  
CPAMID-11



### Anterior Lateral

CPALAT-8



### Drill Guide

GUIDELCBS-2.0  
GUIDELCBS-2.3




### SHIELD

SHIELD



### Cortical Screw

HEX2.7-xx  
HEX3.2-xx  
08mm to 28mm



### Cortical Locking Screw

LCBS2.7-xx  
LCBS3.2-xx  
08mm to 24mm



### Plate Benders

BNDPLT-CFS



TriMed, Inc. / 27533 Avenue Hopkins / Santa Clarita, CA 91355 USA / 800-633-7221 / [www.trimedortho.com](http://www.trimedortho.com)

The technique presented is one suggested surgical technique. The decision to use a specific implant and the surgical technique must be based on sound medical judgment by the surgeon that takes into consideration factors such as the circumstances and configuration of the injury.

This document is controlled by TriMed, Inc. When downloaded, printed, and/or copied, this document becomes uncontrolled, and users should always check [trimedortho.com](http://trimedortho.com) for the latest version.

For indications, contraindications, warnings and precautions related to TriMed Clavicle Fixation System reference IFU on [trimedortho.com/ifu](http://trimedortho.com/ifu).

See [trimedortho.com/patents](http://trimedortho.com/patents) for all patent information.