

Dorsal Hook Plate™


••• Wrist 3





Exposure and Reduction


- Expose the dorsal surface of the distal radius, transposing the EPL tendon, if necessary.
- Reduce the fracture and temporarily fix with 1.1 mm (0.045") K-wires as needed.
- Apply the Dorsal Hook Drill Guide in the desired position.



Stabilize Guide and Check Position


- Insert a 1.1 mm K-wire through the central distal hole of the guide to verify hook positions.
- A second K-wire through the sleeve in the center of the guide can be used to verify position of the locking peg.
- Confirm the K-wire positions with AP and 10° lateral X-ray.

Note: Inserting a 1.1 mm K-wire in either sleeve at the proximal end of the guide can help stabilize the drill guide and help guide the plate into position during insertion.




Drill Distal Holes

- Remove the K-wire used to confirm the locking peg trajectory (if placed) from the sleeve in the center of the guide.
- Cut the distal K-wire 3cm above the surface of the guide to avoid interference with drilling the outer holes.
- Drill the two outer holes at the distal end of the guide with the 1.8mm (blue) drill to penetrate the outer cortex. Slide the guide off the K-wire(s).



Insert Dorsal Hook Plate™

- Assemble the Dorsal Hook Plate™ onto the Hook Plate Inserter. Ensure that the slide is facing away from the implant.
- Pass the distal K-wire into the cannulation in the Hook Plate Inserter handle and engage the hooks into the predrilled hole.
- Impact lightly to fully seat the hooks and plate flush against the bone.
- Remove the Hook Plate Inserter and K-wire(s).



Proximal Fixation

- Use the 1.8mm drill guide (GUIDE-1.8/2.4) and 1.8mm (blue) drill bit for preparing a screw in the proximal slotted hole. Measure length with the depth gauge (GAUGE-1.8).
- Secure plate proximally with 2.4mm cortical screw.
- Use the Locking Peg Guide labeled "FIXED" with the 1.8mm (blue) drill bit for placing locking pegs. Measure length with the markings on the guide or with the depth gauge (GAUGE-1.8).

Note:


1. Use the standard 1.8mm Drill Guide (GUIDE-1.8/2.4) for 2.4mm cortical screws.
2. The Locking Peg Guide labeled "FIXED" is used for both 2.4mm partially threaded (UPEG) and 2.4mm fully threaded (TPEG) locking pegs.



Complete Fixation

- Complete proximal fixation with 2.4mm cortical screws and/or 2.4mm locking pegs as needed.

All components are **Wrist Fixation System 3 (WS3)** items. All implants made from surgical grade stainless steel.

Screw Table	 Unthreaded Peg, 2.4mm	 Threaded Peg 2.4mm	 Cortical Screw, 2.4mm
	UPEG2.4-XX	TPEG2.4-XX	TRX2.4-XX
Length	10-28mm *	10-32mm *	10-32mm *
Drill		● 1.8mm	
Guide		GUIDELF-1.8 GDMINI-1.8	GUIDE-1.8/2.4
Driver		Torx 8	

* 2mm increments ** 1mm increments

Dorsal Hook Plate™

WHKD-4
WHKD-6



Please Note:
Markings on the underside of plate are unique to Wrist 3 Plates *only*.

Dorsal Hook Drill Guide

GDWHKD-1.8



Hook Inserter

INSRTR-WFS



Drill Guide

GUIDE-1.8/2.4
GUIDELF-1.8



K-Wire

WIRE-1.1/100
WIRE-1.6/100



TriMed, Inc. / 27533 Avenue Hopkins / Santa Clarita, CA 91355 USA / 800-633-7221 / www.trimedortho.com

The technique presented is one suggested surgical technique. The decision to use a specific implant and the surgical technique must be based on sound medical judgment by the surgeon that takes into consideration factors such as the circumstances and configuration of the injury.

This document is controlled by TriMed, Inc. When downloaded, printed, and/or copied, this document becomes uncontrolled, and users should always check trimedortho.com for the latest version.

For indications, contraindications, warnings and precautions related to TriMed Wrist Fixation System 3 reference IFU on trimedortho.com/ifu.

See trimedortho.com/patents for all patent information.

